

## PCR-/NAT EHEC / STEC (RV 534) November 2015



**Tabelle 1:** Probenzusammensetzung und erwartetes Ergebnis.

*Sample composition and expected results.*

	Erwartet / expected		Probenzusammensetzung / Sample composition
1525341	Ø	62	ETEC (~5x10 <sup>5</sup> CFU/mL) (ST und LT positive)
1525342	+++	61 / 72	EHEC (~1x10 <sup>5</sup> CFU/mL) (stx-2 positive)
1525343	+++	61 / 72,77,78	EHEC (~5x10 <sup>4</sup> CFU/mL) (stx-2, eae, hlyA and O157 positive)
1525344	Ø	62	<i>Escherichia coli</i> K12 (negative for eae and hlyA)

**Tabelle 2:** Häufigkeit der Mitteilung verschiedener Befunde.

*Absolute numbers of reported individual results.*

n = 124	Probennummer (Sample no.)			
Befund Result	1525341	1525342	1525343	1525344
Positiv	2	118 <sup>1)</sup>	117 <sup>1)</sup>	3
Negativ	122	6	7	121
Fraglich Questionable	0	0	0	0

	Inhibition			
	1525341	1525342	1525343	1525344
n.d.	1	1	1	1
nein no	123	123	123	123
ja yes	0	0	0	0

**Tabelle 3:** Häufigkeit richtig positiver und richtig negativer NAT-Befunde bei Anwendern verschiedener Methoden.

*Absolute numbers and relative frequency of reported true positive and true negative results among various NAT methods.*

NAT-Methode [Code] (total number *)	NAT richtig positiv True positive results			NAT richtig negativ True negative results		
	Absolut Absolute	Relativ Relative	%	Absolut Absolute	Relativ Relative	%
Hain GenoType EHEC [20] (n = 25)	46	46 / 50	92	46	46 / 50	92
Hyplex EHEC [21] (n = 3)	6	6 / 6	100	6	6 / 6	100
r-Biopharm RIDAGENE [22] (n = 35)	67	67 / 70	96	70	70 / 70	100
Other commercial tests [27] (n = 13)	25	25 / 26	96	26	26 / 26	100
In house PCR assay [28] (n = 48)	92	92 / 96	96	96	96 / 96	100
Andere/ k.A. / other [29] (n = 2)	3	3 / 4	75	3	3 / 4	75

\* Durch Mehrfachnennung oder fehlende Angabe kann die absolute Zahl der Ergebnisse (Tab. 2) von der Anzahl der Teilnehmer abweichen.  
Due to reporting results of multiple assay systems or missing specifications, the effective numbers are not correlating with the number of participants

**Comments:** <sup>1)</sup> Partial or complete shiga toxin subtyping, eae-, and hlyA-detection was performed by 104 laboratories. With the exception of 1 laboratory, all reported results were correct.