

## PCR-/NAT EHEC / STEC (RV 534) Mai / Juni 2018



**Tabelle 1:** Probenzusammensetzung und erwartetes Ergebnis.

Sample composition and expected results.

	Erwartet / expected		Probenzusammensetzung / Sample composition
1815341	Ø	62 / 77	EPEC (~1x10 <sup>4</sup> CFU/mL)
1815342	Ø	62	Escherichia coli K12 (negative for eae and hlyA)
1815343	++++	61 / 71,72,77,78	EHEC (~1x10 <sup>6</sup> CFU/mL) (stx-1, stx-2, eae, hlyA and O157 positive)
1815344	+++	61 / 71,72,77,78	EHEC (~1x10 <sup>5</sup> CFU/mL) (stx-1, stx-2, eae, hlyA and O157 positive)

**Tabelle 2:** Häufigkeit der Mitteilung verschiedener Befunde.

Absolute numbers of reported individual results.

n = 140	Probennummer (Sample no.)			
Befund Result	1815341	1815342	1815343	1815344
Positiv	1	0	140 <sup>1)</sup>	140 <sup>1)</sup>
Negativ	139	140	0	0
Fraglich Questionable	0	0	0	0

	Inhibition			
	1815341	1815342	1815343	1815344
n.d.	1	1	1	1
nein no	139	139	139	139
ja yes	0	0	0	0

**Tabelle 3:** Häufigkeit richtig positiver und richtig negativer NAT-Befunde bei Anwendern verschiedener Methoden.

Absolute numbers and relative frequency of reported true positive and true negative results among various NAT methods.

NAT-Methode [Code] (total number *)	NAT richtig positiv True positive results			NAT richtig negativ True negative results		
	Absolut Absolute	Relativ Relative	%	Absolut Absolute	Relativ Relative	%
Hain GenoType EHEC [20] (n = 22)	44	44 / 44	100	44	44 / 44	100
r-Biopharm RIDAGENE [22] (n = 41)	82	82 / 82	100	81	81 / 82	99
Other commercial tests [27] (n = 30)	60	60 / 60	100	60	60 / 60	100
In house PCR assay [28] (n = 50)	100	100 / 100	100	100	100 / 100	100
Andere/ k.A. / other [29] (n = 1)	2	2 / 2	100	2	2 / 2	100

\* Durch Mehrfachnennung oder fehlende Angabe kann die absolute Zahl der Ergebnisse (Tab. 2) von der Anzahl der Teilnehmer abweichen.  
Due to reporting results of multiple assay systems or missing specifications, the effective numbers are not correlating with the number of participants

**Comments:** <sup>1)</sup> Partial or complete shiga toxin subtyping, eae-, and hlyA-detection was performed by 120 laboratories. With the exception of 2 laboratories, all reported results were correct.